

Corrigendum to Bid Ref. No:536/ NCD/ Kpt dtd.17-11-2018.

P. No	Clause	Present Clause	Modified Clause
21	1	Digitizer(CR) system should have the capacity to process more than 60 or more cassette/film per hour of 14 inch x17 inch size.	Digitizer(CR) system should have the capacity to process more than 30 or more cassette/film per hour of 14 inch x 17 inch size.
22	11	Acceptable film size : 14 inch x17 inch , 10 inch x12 inch, 8 inch x10 inch and mammo film size 15 inch x30 inch	Acceptable film size : 14 inch x17 inch , 10 inch x12 inch, 8 inch x10 inch
22	12: Note	The cost of one pack consisting of 100 films from each size of films (4 sizes 14 inch x17 inch , 10 inch x12 inch, 8 inch x10 inch and 15 inch x 30 inch) must be quoted separately in the separate price bid format in PDF(Format A) which shall be taken in to account for evaluation and shall be approved for warranty period of 5 years.	The cost of one pack (<i>mention No. of films in the pack</i>) from each size of 14 inch x17 inch , 10 inch x12 inch and 8 inch x10 inch with warranty period of 5 years must be quoted in price bid form (Annexure-IX) in separate lines under authorised signature in PDF, which shall be taken in to account for evaluation. and shall be approved for The unit price will be for a single film for the purpose of costing and comparison but in case the packing is too big for the user it may influence the whole procure- ment since they are compatible with the machine.
23	X Ray Tube	Anode heat storage capacity of the tube should be more than 140 KHU	Anode heat storage capacity of the tube should be 140 KHU or more.
24	X Ray Tube Head	Mono block version X-ray Tube Head with Rotating Anode single focus X-Ray Tube. The mono block consists of Tube H.V. Transformer, Filament transformer, H.V Rectifiers and Capacitors all immersed in High Grade dielectric oil	Mono block version X-ray Tube Head with Rotating Anode X-Ray Tube. The mono block consists of Tube H.V. Transformer, Filament transformer, H.V Rectifiers and Capacitors all immersed in High Grade dielectric oil
47	Under taking	I/we undertake to get the equipments repaired within 48 hours of receiving of the complaint from the indenting hospital/ consignee failing which a penalty @ 1% of the cost may be recovered from the performance	I/ we undertake to get the equipments repaired within 48 hours of receiving of the complaint from the indenting hospital/ consignee failing which a penalty @ 1% of the cost may be recovered from the performance security before

		security before releasing the same to us after 5 years.	releasing the same to us after 5 years. In exceptional and/or reasonable cases, to the satisfaction of the customer, the above may be relaxed with a condition that the period of breakdown will be added to the warranty period.
5	Delivery period (Section II)	Within 30 days from the issue of the supply order	The supply should be within 30 days from the date of issue of the Purchase order or the AERB permission whichever is later.


 26/11/18
 CDM & PHO Koraput